

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

W3T262906

OGM PLUS 00112VCFPVKE-G1/8i-99-2600p-P

materials of construction: PVC, FKM HF, PVDF, ceramic
signal output (standard): 0,38 ml/pulse, 2600 pulses/l
max. pump size: 11,2 / 13,5 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 01 / 701860 / 23622 / 1001

260162

ECOLAB Engineering GmbH
D-83313 Siegsdorf

Anlaufgrenze 08 [l/h]*
lower measuring range limit

Impulszahl 55 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 13.09.22

Prüfer *P. B. B.* P271

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

W3T262906

OGM PLUS 00112VCFPPVKE-G1/8I-99-2600p-P

materials of construction: PVC, FKM HF, PVDF, ceramic

signal output (standard): 0,38 ml/pulse, 2600 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01 / 701860 / 23622 / 1002

ECOLAB Engineering GmbH
D-83313 Siegsdorf



260162



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 57 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 13.09.2022

Prüfer P271

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

W3T262906

OGM PLUS 00112VCFPPVKE-G1/8i-99-2600p-P

materials of construction:	PVC, FKM HF, PVDF, ceramic
signal output (standard):	0,38 ml/pulse, 2600 pulses/l
max. pump size:	11,2 / 13,5 l/h (50/60-Hz)
semiconductor junction:	PNP
prod. code:	01 / 701860 / 23622 / 1003

ECOLAB Engineering GmbH
D-83313 Stegedorf



260162



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl _____ [n/min]
pulse rate

bei 7,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 14.09.2022

Prüfer P271

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

W3T262906

OGM PLUS 00112VCFPPVKE-G1/8i-99-2600p-P

materials of construction: PVC, FKM HF, PVDF, ceramic

signal output (standard): 0,38 ml/pulse, 2600 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01 / 701860 / 23622 / 1004

ECOLAB Engineering GmbH
D-63313 Sieglar



260162



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 54 [n/min]
pulse rate

bei 7,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 14.05.2022

Prüfer P271

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

W3T262906

OGM PLUS 00112VCFPPVKE-G1/8i-99-2800p-P

materials of construction:	PVC, FKM HF, PVDF, ceramic
signal output (standard):	0,38 ml/pulse, 2800 pulses/l
max. pump size:	11,2 / 13,5 l/h (50/60Hz)
semiconductor junction:	PNP
prod. code:	01 / 701860 / 23622 / 1005

ECOLAB Engineering GmbH
D-83313 Siegsdorf



280162



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 57 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 14.08.2022

Prüfer P271

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

W3T262906

OGM PLUS 00112VCFFPPVKE-G1/BI-99-2800p-P

materials of construction: PVC, FKM HF, PVDF, ceramic
signal output (standard): 0.38 ml/pulse, 2600 pulses/l
max. pump size: 11,2 / 13,5 l/h (50/60Hz)
semiconductor junction: PNP

prod. code: 01 / 701860 / 23622 / 1006

ECOLAB Engineering GmbH
D-63313 Siegsdorf



280162



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 52 [n/min]
pulse rate

bei 7,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 14.09.2022

Prüfer P271

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

W3T262906

OGM PLUS 00112VCFPPVKE-G1/81-99-2600p-P

materials of construction: PVC, FKM HF, PVDF, ceramic

signal output (standard): 0,36 ml/pulse, 2600 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01 / 701860 / 23622 / 1007

ECOLAB Engineering GmbH
D-83313 Siegsdorf



280162

Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 52 [n/min]
pulse rate

bei 7,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 14.09.2022

Prüfer P271

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

W3T262906

OGM PLUS 00112VCFPPVKE-G1/8i-99-2600p-P

materials of construction: PVC, FKM HF, PVDF, ceramic
signal output (standard): 0,38 ml/pulse, 2600 pulses/l
max. pump size: 11,2 / 13,5 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 01 / 701860 / 23622 / 1008

ECOLAB Engineering GmbH
D-83313 Siegsdorf



260162



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 52 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 14.09.2022

Prüfer P271

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

W3T262906

OGM PLUS 00112VCFPPVKE-G1/8i-99-2600p-P

materials of construction: PVC, FKM HF, PVDF, ceramic

signal output (standard): 0,38 ml/pulse, 2600 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01 / 701860 / 23622 / 1009

ECOLAB Engineering GmbH
D-83313 Siegsdorf



280162



Anlaufgrenze 0,5 [l/h]*
lower measuring range limit

Impulszahl 50 [n/min]
pulse rate

bei 7,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 14.09.2022

Prüfer P271

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

W3T262906

OGM PLUS 00112VCFPPVKE-G1/81-99-2600p-P

materials of construction: PVC, FKM HF, PVDF, ceramic

signal output (standard): 0,38 ml/pulse, 2800 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01 / 701860 / 23622 / 1010

ECOLAB Engineering GmbH
D-83313 Siegsdorf



280162



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 50 [n/min]
pulse rate

bei 7,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 14.09.2022

Prüfer P271

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

W3T262906

OGM PLUS 00112VCFPPVKE-G1/8i-99-2800p-P

materials of construction: PVC, FKM HF, PVDF, ceramic

signal output (standard): 0,38 ml/pulse, 2800 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60-Hz)

semiconductor junction: PNP

prod. code: 01 / 701860 / 23622 / 1011

ECOLAB Engineering GmbH
D-83313 Siegsdorf



280162



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 52 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 14.09.2022

Prüfer P271

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

W3T262906

OGM PLUS 00112VCFPPVKE-G1/8i-99-2600p-P

materials of construction:	PVC, FKM HF, PVDF, ceramic
signal output (standard):	0,38 ml/pulse, 2600 pulses/l
max. pump size:	11,2 / 13,5 l/h (50/60Hz)
semiconductor junction:	PNP
prod. code:	01 / 701860 / 23622 / 1012

ECOLAB Engineering GmbH
D-83313 Siegsdorf



280162



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 57 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 14.03.2022

Prüfer P271

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

W3T262906

OGM PLUS 00112VCFPPVKE-G1/8i-99-2600p-P

materials of construction: PVC, FKM HF, PVDF, ceramic

signal output (standard): 0,36 ml/pulse, 2600 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60-Hz)

semiconductor junction: PNP

prod. code: 01 / 701660 / 23622 / 1013

ECOLAB Engineering GmbH
D-83313 Siegsdorf



260162



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 52 [n/min]
pulse rate

bei 7,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 14.08.2012

Prüfer P271

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

W3T262906

OGM PLUS 00112VCFPPVKE-G1/6i-99-2600p-P

materials of construction: PVC, FKM HF, PVDF, ceramic

signal output (standard): 0.36 ml/pulse, 2600 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01 / 701860 / 23622 / 1014

ECOLAB Engineering GmbH
D-83313 Siegsdorf



280162



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 52 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 14.09.2022

Prüfer P271

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

W3T262906

OGM PLUS 00112VCFPPVKE-G1/BI-99-2600p-P

materials of construction: PVC, FKM HF, PVDF, ceramic

signal output (standard): 0,39 ml/pulse, 2600 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01 / 701860 / 23622 / 1015

ECOLAB Engineering GmbH
D-83313 Stegedorf



280162



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 52 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 14.09.2022

Prüfer **P271**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

W3T262906

OGM PLUS 00112VCFPPVKE-G1/8i-99-2800p-P

materials of construction: PVC, FKM HF, PVDF, ceramic
signal output (standard): 0,38 ml/pulse, 2800 pulses/l
max. pump size: 11,2 / 13,5 l/h (50/60-Hz)
semiconductor junction: PNP
prod. code: 01 / 701860 / 23622 / 1016

ECOLAB Engineering GmbH
D-83313 Siegsdorf



280162



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 52 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 14.09.2022

Prüfer P271

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

W3T262906

OGM PLUS 00112VCFPPVKE-G1/8i-99-2600p-P

materials of construction: PVC, FKM HF, PVDF, ceramic

signal output (standard): 0,38 ml/pulse, 2600 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01 / 701880 / 23822 / 1017

ECOLAB Engineering GmbH
D-83313 Siegsdorf



280162



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 52 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 14.03.2022

Prüfer P271

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

W3T262906

OGM PLUS 00112VCFPPVKE-G1/8i-99-2600p-P

materials of construction: PVC, FKM HF, PVDF, ceramic

signal output (standard): 0,38 ml/pulse, 2600 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01 / 701860 / 23622 / 1018

280162

ECOLAB Engineering GmbH
D-83313 Siegedorf

Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 52 [n/min]
pulse rate

bei 7,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 14.09.2022

Prüfer P271

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

W3T262906

OGM PLUS 00112VCFPPVKE-G1/81-99-2600p-P

materials of construction: PVC, FKM HF, PVDF, ceramic
signal output (standard): 0,38 ml/pulse, 2600 pulses/l
max. pump size: 11,2 / 13,5 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 01 / 701860 / 23622 / 1019

ECOLAB Engineering GmbH
D-83313 Siegsdorf



280162



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 50 [n/min]
pulse rate

bei 7,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 14.09.2022

Prüfer **P271**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112



W3T262906

OGM PLUS 00112VCFPPVKE-G1/8i-99-2600p-P

materials of construction: PVC, FKM HF, PVDF, ceramic
signal output (standard): 0,38 ml/pulse, 2600 pulses/l
max. pump size: 11,2 / 13,5 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 01 / 701860 / 23622 / 1020

260162

ECOLAB Engineering GmbH
D-63313 Siesdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 52 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 14.09.2022

Prüfer P271

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

W3T262906

OGM PLUS 00112VCFPPVKE-G1/BI-99-2800p-P

materials of construction:	PVC, FKM HF, PVDF, ceramic
signal output (standard):	0,38 ml/pulse, 2800 pulses/l
max. pump size:	11,2 / 13,5 l/h (50/60-Hz)
semiconductor junction:	PNP
prod. code:	01 / 701860 / 23622 / 1021

ECOLAB Engineering GmbH
D-83313 Siegsdorf



280162



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 52 [n/min]
pulse rate

bei 7,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 14.09.2022

Prüfer P271

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

W3T262906

OGM PLUS 00112VCFPPVKE-G1/81-99-2600p-P

materials of construction:	PVC, FKM HF, PVDF, ceramic
signal output (standard):	0,38 ml/pulse, 2600 pulses/l
max. pump size:	11,2 / 13,5 l/h (50/60Hz)
semiconductor junction:	PNP
prod. code:	01 / 701860 / 23622 / 1022

ECOLAB Engineering GmbH
D-83313 Siegsdorf



280162

Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 51 [n/min]
pulse rate

bei 7,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 19.09.2022

Prüfer P271

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112



W3T262906

OGM PLUS 00112VCFPPVKE-G1/8i-99-2600p-P

materials of construction: PVC, FKM HF, PVDF, ceramic
signal output (standard): 0,39 ml/pulse, 2600 pulses/l
max. pump size: 11,2 / 13,5 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 01 / 701860 / 23622 / 1023

280162

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 54 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 14.09.2022

Prüfer **P271**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

W3T262906

OGM PLUS 00112VCFPPVKE-G1/6i-99-2600p-P

materials of construction: PVC, FKM HF, PVDF, ceramic

signal output (standard): 0,36 ml/pulse, 2600 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60-Hz)

semiconductor junction: PNP

prod. code: 01 / 701880 / 23622 / 1024

ECOLAB Engineering GmbH
D-83313 Siegsdorf



280162



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 54 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 14.9.22

Prüfer P271

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

W3T262906

OGM PLUS 00112VCFPPVKE-G1/6i-99-2600p-P

materials of construction: PVC, FKM HF, PVDF, ceramic

signal output (standard): 0,38 ml/pulse, 2600 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01 / 701860 / 23622 / 1025

ECOLAB Engineering GmbH
D-63313 Siegedorf



280162



Anlaufgrenze 0,7 [l/h]*
lower measuring range limit

Impulszahl 52 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 14.09.2022

Prüfer P271

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

W3T262906

OGM PLUS 00112VCFPPVKE-G1/BI-99-2600p-P

materials of construction: PVC, FKM HF, PVDF, ceramic
signal output (standard): 0,38 ml/pulse, 2600 pulses/l
max. pump size: 11,2 / 13,5 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 01 / 701860 / 23622 / 1026

ECOLAB Engineering GmbH
D-83313 Siegsdorf



280162



Anlaufgrenze 0,7 [l/h]*
lower measuring range limit

Impulszahl 55 [n/min]
pulse rate

bei 7,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 14.03.2022

Prüfer P271

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

W3T262906

OGM PLUS 00112VCFPPVKE-G1/8i-99-2600p-P

materials of construction: PVC, FKM HF, PVDF, ceramic

signal output (standard): 0,38 ml/pulse, 2600 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01 / 701860 / 23622 / 1027

ECOLAB Engineering GmbH
D-83313 Siegedorf



280162



Anlaufgrenze 0,6 [l/h]*
lower measuring range limit

Impulszahl 53 [n/min]
pulse rate

bei 7,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 14.09.2022

Prüfer P271

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

W3T262906

OGM PLUS 00112VCFPPVKE-G1/81-99-2600p-P

materials of construction:	PVC, FKM HF, PVDF, ceramic
signal output (standard):	0,38 ml/pulse, 2600 pulses/l
max. pump size:	11,2 / 13,5 l/h (50/60Hz)
semiconductor junction:	PNP
prod. code:	01 / 701860 / 23622 / 1028

ECOLAB Engineering GmbH
D-83313 Siegsdorf



280162

Anlaufgrenze 0,5 [l/h]*
lower measuring range limit

Impulszahl 57 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 14.09.2012

Prüfer P271

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

W3T262908

OGM PLUS 00112VCFPPVKE-G1/81-99-2600p-P

materials of construction: PVC, FKM HF, PVDF, ceramic
signal output (standard): 0,38 ml/pulse, 2600 pulses/l
max. pump size: 11,2 / 13,5 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 01 / 701880 / 23622 / 1029

ECOLAB Engineering GmbH
D-83313 Siegedorf



280162



Anlaufgrenze 0,6 [l/h]*
lower measuring range limit

Impulszahl 51 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 15.09.2022

Prüfer **P271**