

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00540	
PN: 280073	
OGM PLUS 00540VCEPPKKE-G1/4I-99-0005m-P	
materials of construction:	PVC, EPDM, PEEK, ceramic
signal output (standard):	5,0 ml/pulse, 200 pulses/l
max. pump size:	54 / 65 l/h (50/60Hz)
semiconductor junction:	PNP
prod. code:	01 / 704522 / 13722 / 1001
ECOLAB Engineering GmbH D-83313 Siegedorf	
	

Anlaufgrenze 6 [l/h]*
lower measuring range limit

Impulszahl 33 [n/min]
pulse rate

bei 10 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 20.09.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00540

PN: **280073**



OGM PLUS 00540VCEPPKKE-G1/4I-99-0005m-P

materials of construction: PVC, EPDM, PEEK, ceramic

signal output (standard): 5.0 ml/pulse, 200 pulses/l

max. pump size: 54 / 85 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01 / 704522 / 13722 / 1002

ECOLAB Engineering GmbH
D-63313 Siegedorf



Anlaufgrenze 6 [l/h]*
lower measuring range limit

Impulszahl 33 [n/min]
pulse rate



bei 10 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 20.09.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00540	
PN: 280073	
OGM PLUS 00540VCEPPKKE-G1/4I-99-0005m-P	
materials of construction:	PVC, EPDM, PEEK, ceramic
signal output (standard):	5,0 ml/pulse, 200 pulses/l
max. pump size:	54 / 65 l/h (50/60Hz)
semiconductor junction:	PNP
prod. code:	01/704522/13722/1003
ECOLAB Engineering GmbH D-83313 Siegsdorf	

Anlaufgrenze 6 [l/h]*
lower measuring range limit

Impulszahl 35 [n/min]
pulse rate

bei 10 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 20.09.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00540		
PN: 280073		
OGM PLUS 00540VCEPPKKE-G1/4I-99-0005m-P		
materials of construction:	PVC, EPDM, PEEK, ceramic	
signal output (standard):	5.0 ml/pulse, 200 pulses/l	
max. pump size:	54 / 65 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01/704522/10722/1004	
ECOLAB Engineering GmbH D-83313 Siegsdorf		

Anlaufgrenze 6 [l/h]*
lower measuring range limit

Impulszahl 34 [n/min]
pulse rate

bei 10 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 20.09.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00540
PN: **280073**
OGM PLUS 00540VCEPPKKE-G1/4I-99-0005m-P
materials of construction: PVC, EPDM, PEEK, ceramic
signal output (standard): 5,0 ml/pulse, 200 pulses/l
max. pump size: 54 / 85 l/h (50/60Hz)
semiconductor junction: PNP
prod. code: 01 / 704522 / 13722 / 1005

ECOLAB Engineering GmbH
D-83313 Siegedorf

Anlaufgrenze 6 [l/h]*
lower measuring range limit

Impulszahl 36 [n/min]
pulse rate

bei 10 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 20.09.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00540

PN: **280073**

OGM PLUS 00540VCEPPKKE-G1/4I-99-0005m-P

materials of construction: PVC, EPDM, PEEK, ceramic

signal output (standard): 5.0 ml/pulse, 200 pulses/l

max. pump size: 54 / 65 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01 / 704522 / 13722 / 1008

ECOLAB Engineering GmbH
D-63313 Siegedorf



Anlaufgrenze 6 [l/h]*
lower measuring range limit

Impulszahl 32 [n/min]
pulse rate

bei 10 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 20.09.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00540

PN: **280073**

OGM PLUS 00540VCEPPKKE-G1/4I-99-0005m-P

materials of construction: PVC, EPDM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 54 / 65 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01 / 704522 / 13722 / 1007

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 6 [l/h]*
lower measuring range limit

Impulszahl 34 [n/min]
pulse rate

bei 10 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 20.09.22

Prüfer **P 410**

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00540		
PN: 280073		
OGM PLUS 00540VCEPPKKE-G1/4I-99-0005m-P		
materials of construction:	PVC, EPDM, PEEK, ceramic	
signal output (standard):	5.0 ml/pulse, 200 pulses/l	
max. pump size:	54 / 65 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01 / 704522 / 13722 / 1008	
ECOLAB Engineering GmbH D-83313 Siegsdorf		

Anlaufgrenze 6 [l/h]*
lower measuring range limit

Impulszahl 32 [n/min]
pulse rate

bei 10 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 20.09.22

Prüfer P_410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00540

PN: **280073**

OGM PLUS 00540VCEPPKKE-G1/4I-99-0005m-P

materials of construction: PVC, EPDM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 54 / 65 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01 / 704522 / 13722 / 1009

ECOLAB Engineering GmbH
D-83313 Sieddorf



Anlaufgrenze 6 [l/h]*
lower measuring range limit

Impulszahl 33 [n/min]
pulse rate

bei 10 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 20.09.22

Prüfer P 410

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00540

PN: **280073**

OGM PLUS 00540VCEPPKKE-G1/4I-99-0005m-P

materials of construction: PVC, EPDM, PEEK, ceramic

signal output (standard): 5.0 ml/pulse, 200 pulses/l

max. pump size: 54 / 65 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/704522/13722/1010

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 6 [l/h]*
lower measuring range limit

Impulszahl 33 [n/min]
pulse rate

bei 10 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 20.09.22

Prüfer P 410