


Prüfprotokoll OGM^{plus}

Oval gear meter type 4		
PN: 280004		
OGM 040VCFPKKETI-G3/4i-99-0300-0120		
Materials of construction:	PVC, FPM, PEEK, ceramic, titan	
Signal output:	300 pulse/l	
Max. flow rate:	120 l/h	
Prod. code:	702328 / 231 22 / 1001	
ECOLAB Engineering GmbH D-83313 Siegsdorf		

Anlaufgrenze 20 [l/h]*
lower measuring range limit

Impulszahl 166 [n/min]
pulse rate

bei 35 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 03.08.22

Prüfer **P 410**