

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280140**

OGM PLUS 00112PVFPPVKE-G1/8I-99-00005m-P

materials of construction: PVDF, FPM, PVDF, ceramic

signal output (standard): 0,5 ml/pulse, 2000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/703629/15022/1001

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 42 [n/min]
pulse rate



bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 14.12.22

Prüfer Sollnereder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112		
PN: 280140		
OGM PLUS 00112PVFPPVKE-G1/81-99-00005m-P		
materials of construction:	PVDF, FPM, PVDF, ceramic	
signal output (standard):	0,5 ml/pulse, 2000 pulses/l	
max. pump size:	11,2 / 13,5 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01/703629/15022/1002	
ECOLAB Engineering GmbH		
D-83313 Siegedorf		

Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 61 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 15.12.22

Prüfer Sollnereder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280140**

OGM PLUS 00112PVFPPVKE-G1/8I-99-00005m-P

materials of construction: PVDF, FPM, PVDF, ceramic

signal output (standard): 0,5 ml/pulse, 2000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/703929/15022/1003

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 40 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 15.12.22

Prüfer Sollereder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280140**

OGM PLUS 00112PVFPPVKE-G1/8I-99-00005m-P

materials of construction: PVDF, FPM, PVDF, ceramic

signal output (standard): 0,5 ml/pulse, 2000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/703629/15022/1004

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 42 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 15.12.22

Prüfer Solleredes

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280140**

OGM PLUS 00112PVFPPVKE-G1/8I-99-00005m-P

materials of construction: PVDF, FPM, PVDF, ceramic

signal output (standard): 0,5 ml/pulse, 2000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/703829/15022/1005

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 47 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 15.12.22

Prüfer Soller edes

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280140**

OGM PLUS 00112PVFPPVKE-G1/BI-99-00005m-P

materials of construction: PVDF, FPM, PVDF, ceramic

signal output (standard): 0,5 ml/pulse, 2000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/703629/15022/1006

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 41 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 15.12.22

Prüfer Sollerseder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112		
PN: 280140		
OGM PLUS 00112PVFPPVKE-G1/8I-99-00005m-P		
materials of construction:	PVDF, FPM, PVDF, ceramic	
signal output (standard):	0,5 ml/pulse, 2000 pulses/l	
max. pump size:	11,2 / 13,5 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	01/703629/15022/1007	
ECOLAB Engineering GmbH D-83313 Siegsdorf		

Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 39 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 15.12.22

Prüfer Sollers edler

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280140**

OGM PLUS 00112PVFPPVKE-G1/8I-99-00005m-P

materials of construction: PVDF, FPM, PVDF, ceramic

signal output (standard): 0,5 ml/pulse, 2000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60-Hz)

semiconductor junction: PNP

prod. code: 01/703829/15022/1008

ECOLAB Engineering GmbH
D-83313 Slegedorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 39 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 15.12.22

Prüfer Sollereder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **280140**

OGM PLUS 00112PVFPPVKE-G1/8I-99-00005m-P

materials of construction: PVDF, FPM, PVDF, ceramic

signal output (standard): 0,5 ml/pulse, 2000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01/703629/15022/1009

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 4,1 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 14.12.22

Prüfer Sollereder

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 00112

PN: **250140**

OGM PLUS 00112PVFPPVKE-G1/8I-99-00005m-P

materials of construction: PVDF, FPM, PVDF, ceramic

signal output (standard): 0,5 ml/pulse, 2000 pulses/l

max. pump size: 11,2 / 13,5 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 01 / 703629 / 15022 / 1010

ECOLAB Engineering GmbH
D-83113 Siegsdorf



Anlaufgrenze 0,8 [l/h]*
lower measuring range limit

Impulszahl 60 [n/min]
pulse rate

bei 1,25 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 15.12.22

Prüfer Sollmeseder