



Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 02100		
PN: 280045		
OGM PLUS 02100VCFPPKKE-G5/4I-99-0020m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	20.0 ml/pulse, 50 pulses/l	
max. pump size:	210 / 250 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	02/704943/23922/1001	
ECOLAB Engineering GmbH D-83313 Siegsdorf		

Anlaufgrenze 50 [l/h]*
lower measuring range limit

Impulszahl 71 [n/min]
pulse rate

bei 80 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.09.22

Prüfer P3

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 02100

PN: **280045**

OGM PLUS 02100VCFPPKKE-G5/4I-99-0020m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 20,0 ml/pulse, 50 pulses/l

max. pump size: 210 / 250 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/704943 / 23922 / 1002

ECOLAB Engineering GmbH
D-83313 Siegsdorf



Anlaufgrenze 50 [l/h]*
lower measuring range limit

Impulszahl 74 [n/min]
pulse rate

bei 80 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.09.22

Prüfer P3

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 02100

PN: **280045**



OGM PLUS 02100VCFPPKE-G5/4I-99-0020m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 20,0 ml/pulse, 50 pulses/l

max. pump size: 210 / 250 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/704943/23922/1003

ECOLAB Engineering GmbH
D-63313 Siegedorf



Anlaufgrenze 50 [l/h]*
lower measuring range limit

Impulszahl 72 [n/min]
pulse rate

bei 80 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.09.22

Prüfer P3

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 02100	
PN: 280045	
OGM PLUS 02100VCFPPKKE-G5/4I-99-0020m-P	
materials of construction:	PVC, FPM, PEEK, ceramic
signal output (standard):	20,0 ml/pulse, 50 pulses/l
max. pump size:	210 / 250 l/h (50/60Hz)
semiconductor junction:	PNP
prod. code:	02/704943 / 23922 / 1004
ECOLAB Engineering GmbH D-83113 Siegedorf	

Anlaufgrenze 50 [l/h]*
lower measuring range limit

Impulszahl 70 [n/min]
pulse rate

bei 80 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.09.22

Prüfer P3

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 02100

PN: **280045**

OGM PLUS 02100VCFPPKKE-G5/4I-99-0020m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 20,0 ml/pulse, 50 pulses/l

max. pump size: 210 / 250 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/704943/23922/1005

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 50 [l/h]*
lower measuring range limit

Impulszahl 73 [n/min]
pulse rate

bei 80 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.09.22

Prüfer P3

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 02100

PN: **280045**



OGM PLUS 02100VCFPPKKE-G5/4l-99-0020m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 20.0 ml/pulse, 50 pulses/l

max. pump size: 210 / 250 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/704943/23922/1006

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 50 [l/h]*
lower measuring range limit

Impulszahl 70 [n/min]
pulse rate

bei 80 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.09.22

Prüfer P3

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 02100

PN: **280045**



OGM PLUS 02100VCFPPKKE-G5/4I-99-0020m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 20,0 ml/pulse, 50 pulses/l

max. pump size: 210 / 250 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/704943 / 23922 / 1007

ECOLAB Engineering GmbH
D-63313 Siegedorf



Anlaufgrenze 50 [l/h]*
lower measuring range limit

Impulszahl 73 [n/min]
pulse rate

bei 80 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.09.22

Prüfer P3

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 02100

PN: **280045**



OGM PLUS 02100VCFPPKKE-G5/4I-99-0020m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 20.0 ml/pulse, 50 pulses/l

max. pump size: 210 / 250 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/704943/23922/1008

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 50 [l/h]*
lower measuring range limit

Impulszahl 74 [n/min]
pulse rate

bei 80 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.09.22

Prüfer P3

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 02100

PN: **280045**



OGM PLUS 02100VCFPPKKE-G5/4I-99-0020m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 20,0 ml/pulse, 50 pulses/l

max. pump size: 210 / 250 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/704943 / 23922 / 1009

ECOLAB Engineering GmbH
D-63313 Siegsdorf



Anlaufgrenze 50 [l/h]*
lower measuring range limit

Impulszahl 71 [n/min]
pulse rate

bei 80 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.09.22

Prüfer P3

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 02100

PN: **280045**

OGM PLUS 02100VCFPPKKE-G5/4I-99-0020m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 20,0 ml/pulse, 50 pulses/l

max. pump size: 210 / 250 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/704943/23922/1010

ECOLAB Engineering GmbH

D-83313 Sieddorf



Anlaufgrenze 50 [l/h]*
lower measuring range limit

Impulszahl 73 [n/min]
pulse rate

bei 80 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.09.22

Prüfer P3

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 02100

PN: **280045**

OGM PLUS 02100VCFPPKKE-G5/4I-99-0020m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 20,0 ml/pulse, 50 pulses/l

max. pump size: 210 / 250 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/704943/23922/1011

ECOLAB Engineering GmbH
D-83313 Sieddorf



Anlaufgrenze 50 [l/h]*
lower measuring range limit

Impulszahl 74 [n/min]
pulse rate

bei 80 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.09.22

Prüfer P3

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 02100

PN: **280045**

OGM PLUS 02100VCFPPKKE-G5/41-99-0020m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 20.0 ml/pulse, 50 pulses/l

max. pump size: 210 / 250 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/704943/23922/1012

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 50 [l/h]*
lower measuring range limit

Impulszahl 70 [n/min]
pulse rate

bei 80 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.09.22

Prüfer P3

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 02100

PN: **260045**

OGM PLUS 02100VCFPPKKE-G5/4I-99-0020m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 20,0 ml/pulse, 50 pulses/l

max. pump size: 210 / 250 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/704943 / 23922 / 1013

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 50 [l/h]*
lower measuring range limit

Impulszahl 68 [n/min]
pulse rate

bei 80 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.09.22

Prüfer P3

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 02100

PN: **280045**



OGM PLUS 02100VCFPPKKE-G5/4I-99-0020m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 20,0 ml/pulse, 50 pulses/l

max. pump size: 210 / 250 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/704943/23922/1014

ECOLAB Engineering GmbH
D-83313 Stegendorf



Anlaufgrenze 50 [l/h]*
lower measuring range limit

Impulszahl 73 [n/min]
pulse rate

bei 80 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.09.22

Prüfer P3

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 02100

PN: **280045**

OGM PLUS 02100VCFPPKKE-G5/4I-99-0020m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 20,0 ml/pulse, 50 pulses/l

max. pump size: 210 / 250 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/704943/23922/1015

ECOLAB Engineering GmbH
D-83313 Stegedorf



Anlaufgrenze 50 [l/h]*
lower measuring range limit

Impulszahl 72 [n/min]
pulse rate

bei 80 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.09.22

Prüfer P3

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 02100

PN: **280045**



OGM PLUS 02100VCFPPKKE-G5/4I-99-0020m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 20,0 ml/pulse, 50 pulses/l

max. pump size: 210 / 250 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/704943/23922/1016

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 50 [l/h]*
lower measuring range limit

Impulszahl 73 [n/min]
pulse rate


bei 80 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.09.22

Prüfer P3

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 02100		
PN: 280045		
OGM PLUS 02100VCFPPKKE-G5/4I-99-0020m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	20,0 ml/pulse, 50 pulses/l	
max. pump size:	210 / 250 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	02/704943 /23922 /1017	
ECOLAB Engineering GmbH D-83113 Siegsdorf		

Anlaufgrenze 50 [l/h]*
lower measuring range limit

Impulszahl 71 [n/min]
pulse rate

bei 80 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.09.22

Prüfer P3

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 02100		
PN: 280045		
OGM PLUS 02100VCFPPKKE-G5/4I-00-0020m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	20,0 ml/pulse, 50 pulses/l	
max. pump size:	210 / 250 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	02/704943/23922/1018	
ECOLAB Engineering GmbH D-63313 Siegedorf		

Anlaufgrenze 50 [l/h]*
lower measuring range limit

Impulszahl 68 [n/min]
pulse rate

bei 80 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.09.22

Prüfer P3

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 02100

PN: **280045**

OGM PLUS 02100VCFPPKKE-G5/4I-99-0020m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 20,0 ml/pulse, 50 pulses/l

max. pump size: 210 / 250 l/h (50/60Hz)

semiconductor junction: PNP

prod. code: 02/704943/23922/1019

ECOLAB Engineering GmbH
D-83313 Siegedorf



Anlaufgrenze 50 [l/h]*
lower measuring range limit

Impulszahl 73 [n/min]
pulse rate

bei 80 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.09.22

Prüfer P3

Prüfprotokoll OGM^{plus}

Oval gear meter PLUS 02100		
PN: 280045		
OGM PLUS 02100VCFPPKKE-G5/41-99-0020m-P		
materials of construction:	PVC, FPM, PEEK, ceramic	
signal output (standard):	20,0 ml/pulse, 50 pulses/l	
max. pump size:	210 / 250 l/h (50/60Hz)	
semiconductor junction:	PNP	
prod. code:	02/704943/23922/1020	
ECOLAB Engineering GmbH D-83313 Siegedorf		

Anlaufgrenze 50 [l/h]*
lower measuring range limit

Impulszahl 71 [n/min]
pulse rate

bei 80 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 28.09.22

Prüfer P3