

Prüfprotokoll OGM

Washtronic EXACT II Typ1

PN: 280032

OGM 010VCFPKKEI-G1/4I-99-0820-0030

Materials of construction:

PVC, FPM, PEEK,

ceramic, titan

Signal output:

820 pulse/l

Max. flow rate:

30 l/h

Prod. code:

706592 / 34922 / 1001

ECOLAB Engineering GmbH

D-83313 Siegedorf



Anlaufgrenze 55 [l/h]*
lower measuring range limit

Impulszahl 121 [n/min]
pulse rate

bei 9,0 [l/h]*
at

* = kontinuierlicher Durchfluss / continuous flow

Datum 8.12.22

Prüfer P391