

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: 10240555



OGM PLUS 01200VCFPPKKE-G3/4I-99-0005m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

Year/Prod. code: 2022 / 707888/547 22 / 1006

ECOLAB Engineering GmbH  
D-83313 Siegsdorf  
Made in Germany



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 108 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 28.11.22

Prüfer Soltes edler

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: **10240555**

OGM PLUS 01200VCFPPKKE-G3/4I-99-0005m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

Year/Prod. code: 2022 / 707866 / 547 22 / 1018

ECOLAB Engineering GmbH  
D-83313 Siegsdorf  
Made in Germany



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 112 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 28.11.22

Prüfer Sollerseder

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: **10240555**

OGM PLUS 01200VCFPPKKE-G3/4I-99-0005m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

Year/Prod. code: 2022 / 707866/547 22 / 1017

ECOLAB Engineering GmbH  
D-83313 Siegedorf  
Made in Germany



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 116 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 28.11.22

Prüfer Sollwender

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: **10240555**



OGM PLUS 01200VCFPPKKE-G3/4i-99-0005m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

Year/Prod. code: 2022 / 707868 / 547 22 / 1018

ECOLAB Engineering GmbH

D-83313 Siegsdorf

Made in Germany



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 119 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 28.11.22

Prüfer Sollereder

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: **10240555**



OGM PLUS 01200VCFPPKKE-G3/4I-99-0005m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

Year/Prod. code: 2022 / 707866 / 547 22 / 1015

ECOLAB Engineering GmbH  
D-83313 Siegedorf  
Made in Germany



Anlaufgrenze 70 [l/h]\*  
lower measuring range limit

Impulszahl 106 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 28.

Prüfer \_\_\_\_\_

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: **10240555**



OGM PLUS 01200VCFPPKKE-G3/4i-99-0005m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

Year/Prod. code: 2022 / 707868 / 547 22 / 1014

ECOLAB Engineering GmbH

D-83313 Siegsdorf

Made in Germany



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 116 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 28.11.22

Prüfer Sollereder

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: 10240555

OGM PLUS 01200VCFPPKKE-G3/4I-99-0005m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 5.0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

Year/Prod. code: 2022 / 707886/547 22 / 1013

ECOLAB Engineering GmbH  
D-83313 Slegsdorf  
Made in Germany



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 115 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 28.11.22

Prüfer Sollerseder

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200



PN: 10240555

OGM PLUS 01200VCFPPKKE-G3/4I-99-0005m-P

materials of construction: PVC, FPM, PEEK, ceramic  
signal output (standard): 5.0 ml/pulse, 200 pulses/l  
max. pump size: 120 / 144 l/h (50/60Hz)  
semiconductor junction: PNP  
Year/Prod. code: 2022 / 707866/547 22/1012

ECOLAB Engineering GmbH  
D-83313 Siegedorf  
Made in Germany



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 115 [n/min]  
pulse rate

bei 85 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 28.11.22

Prüfer Johannes



# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: 10240555

OGM PLUS 01200VCFPPKKE-G3/4I-99-0005m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 5.0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

Year/Prod. code: 2022 / 707866/547 22 / 1011

ECOLAB Engineering GmbH

D-83313 Stegledt

Made in Germany



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 116 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 28.11.22

Prüfer Sollgeder

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: **10240555**



OGM PLUS 01200VCFPPKKE-G3/4I-99-0005m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

Year/Prod. code: 2022 / 707866/54722/1010

ECOLAB Engineering GmbH  
D-83313 Siegsdorf  
Made in Germany



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 115 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 28.11.22

Prüfer Sollereder

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: **10240555**

OGM PLUS 01200VCFPPKKE-G3/4i-09-0005m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 5.0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

Year/Prod. code: 2022 / 707866 / 547 22 / 1009

ECOLAB Engineering GmbH

D-83313 Siegsdorf

Made in Germany



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 110 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 28.11.22

Prüfer Sollweider

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: **10240555**



OGM PLUS 01200VCFPPKKE-G3/4I-99-0005m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 5.0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

Year/Prod. code: 2022 / 707866/547 22 / 1008

ECOLAB Engineering GmbH  
D-83113 Slegsdorf  
Made in Germany



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 108 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 28.11.22

Prüfer Sollweider

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200



PN: 10240555

OGM PLUS 01200VCFPPKKE-G3/4I-99-0005m-P

materials of construction: PVC, FPM, PEEK, ceramic  
signal output (standard): 5.0 ml/pulse, 200 pulses/l  
max. pump size: 120 / 144 l/h (50/60Hz)  
semiconductor junction: PNP  
Year/Prod. code: 2022 / 707868 / 547 22 / 1007

ECOLAB Engineering GmbH  
D-83313 Siedsdorf  
Made in Germany



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 116 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 28-11-22

Prüfer Sollereder

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: **10240555**



OGM PLUS 01200VCFPPKKE-G3/4I-99-0005m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

Year/Prod. code: 2022 / 707888/547 22 / 1005

ECOLAB Engineering GmbH  
D-83313 Siegsdorf  
Made in Germany



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 119 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 28.11.22

Prüfer Sollnereder

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: **10240555**

OGM PLUS 01200VCFPPKKE-G3/4i-99-0005m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

Year/Prod. code: 2022 / 707866/547 22 / 1004

ECOLAB Engineering GmbH

D-83313 Siegsdorf

Made in Germany



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 113 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 28.11.22

Prüfer Sollner

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: **10240555**



OGM PLUS 01200VCFPPKE-G3/4I-99-0005m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 5.0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

Year/Prod. code: 2022 / 707866/547 22 / 1003

ECOLAB Engineering GmbH  
D-83313 Siegsdorf  
Made in Germany



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 114 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 28.11.22

Prüfer S. Kerschler



# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: 10240555



OGM PLUS 01200VCFPPKKE-G3/4I-99-0005m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 5,0 m/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

Year/Prod. code: 2022 / 707868/547 22 / 1002

ECOLAB Engineering GmbH  
D-83313 Slegsdorf  
Made in Germany



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 110 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 28.11.22

Prüfer Soller edler

# Prüfprotokoll OGM<sup>plus</sup>

Oval gear meter PLUS 01200

PN: **10240555**

OGM PLUS 01200VCFPKKE-G3/4I-99-0005m-P

materials of construction: PVC, FPM, PEEK, ceramic

signal output (standard): 5,0 ml/pulse, 200 pulses/l

max. pump size: 120 / 144 l/h (50/60Hz)

semiconductor junction: PNP

Year/Prod. code: 2022 / 707886/547 22 /1001

ECOLAB Engineering GmbH  
D-83313 Siegsdorf  
Made in Germany



Anlaufgrenze 20 [l/h]\*  
lower measuring range limit

Impulszahl 116 [n/min]  
pulse rate

bei 35 [l/h]\*  
at

\* = kontinuierlicher Durchfluss / continuous flow

Datum 28.11.22

Prüfer Sollerseder